КОМПЛЕКСНЫЙ $^2\text{H}/^1\text{H}$ ЯМР-СКРИНИНГ ДЕЙТЕРИЯ В КОМПОНЕНТАХ ВОДНО-ОРГАНИЧЕСКИХ РАСТВОРОВ*

Г.А.Калабин, д.х.н., В.А.Ивлев, Н.А.Комаров, А.Ю.Колеснов, д.т.н., РУДН
kalabinga@mail.ru, irms@biolab.ru

УДК 543.51; ББК 02.00.02

Предложен алгоритм использования совокупности спектроскопии ЯМР ядер протия ^1H и дейтерия $^2\text{H(D)}$ для быстрой идентификации и определения компонентного и дифференциального изотопного состава водно-органических растворов, в том числе этанола (показатели $(\text{D/H})_I$ и $(\text{D/H})_{II}$) и воды в винной и коньячной продукции, соках, безалкогольных напитках и других продуктах. Метод не требует пробоподготовки и использования аттестованных стандартных образцов. Представлены примеры применения предлагаемого подхода для аутентификации и контроля качества винодельческой продукции.

ВВЕДЕНИЕ

Состав любых водно-органических растворов (ВОР), например, алкогольных напитков, молока, вин, соков и др., содержащих элементы H, C, N, O и S может быть охарактеризован на различных иерархических уровнях – компонентном, фрагментном, атомном и изотопном. Комплексный анализ полученных взаимодополняющих данных позволяет идентифицировать компоненты и происхождение ВОР с высокой степенью точности. Наибольшее распространение для этих целей получили методы элементного и компонентного анализа. Последний преимущественно использует различные виды хроматографии, которая в сочетании с масс-спектрометрическим детектированием обладает высокой чувствительностью и при наличии аттестованных стандартных образцов (АСО) дает возможность идентифицировать и количественно измерять содержание макро- и микрокомпонентов. При изучении ВОР неизвестного состава несомненными достоинствами обладает количественная спектроскопия ядерного магнитного резонанса (ЯМР) ^1H и ^{13}C , не требующая использования АСО. Однако даже сочетание этих методов не обеспечивает строгого доказательства соответствия компонентов ВОР заявленному сортовому, географическому (агроклиматическому)

и/или технологическому происхождению, например, вина защищенного географического указания (ВЗГУ) или защищенного наименования места происхождения (ВЗНМП), минеральной воды – источнику, соков – способу их приготовления (прямой отжим или восстановление из концентрированного сырья), молока – цельному, разбавленному или восстановленному. Компонентный состав ВОР, если это экономически оправдано, можно имитировать, подготовив раствор из произвольной чистой воды и всех необходимых и широкодоступных неорганических и органических компонентов. Однако фальсификат даже при тщательном изготовлении и подборе рецептуры не будет соответствовать подлинному продукту по содержанию минорных стабильных изотопов водорода (^2H), углерода (^{13}C), азота (^{15}N), кислорода (^{17}O и ^{18}O) и серы (^{33}S и ^{34}S). Выявить подделку можно лишь методами изотопного анализа.

Ведущее место в изотопном анализе занимает масс-спектрометрия изотопных отношений (МСИО – IRMS/SIRA) благодаря ее высочайшей чувствительности и прецизионности. Первоначальные трудности в применении этого метода были связаны с наиболее сложным этапом – выделением компонентов объекта и их очисткой для последующей аутентификации. Развитие

* Публикация подготовлена при поддержке Программы РУДН "5-100".

инструментария МСИО, например, появление ГХ- (для исследования летучих соединений) и ВЭЖХ-систем (для исследования нелетучих соединений), комбинированных с соответствующими изотопными интерфейсами, позволило исключить трудоемкие операции по извлечению аналитов из матрицы. Как правило, методом МСИО определяют так называемые интегральные отношения изотопов $^2\text{H}/^1\text{H}$, $^{13}\text{C}/^{12}\text{C}$, $^{15}\text{N}/^{14}\text{N}$, $^{18}\text{O}/^{16}\text{O}$, $^{34}\text{S}/^{32}\text{S}$. Например, для молекулы этилового спирта $\text{CH}_3\text{CH}_2\text{OH}$ в составе вина можно измерить только одно значение $^2\text{H}/^1\text{H}$ ($\delta^2\text{H}(\text{D})$) или $^{13}\text{C}/^{12}\text{C}$ ($\delta^{13}\text{C}$). Однако, для этих изотопов характерно неравновесное распределение в отдельных фрагментах молекулы, то есть три и два разных значения для водорода и углерода соответственно. Соотношение между ними отражает специфику агроклиматических условий географического региона выращивания винограда. Кроме того, при исследовании, например, спиртосодержащих дистиллятов необходимо учитывать присутствие воды, которая влияет на достоверность результатов при измерении изотопии водорода ($^2\text{H}(\text{D})/^1\text{H}$) и кислорода ($^{18}\text{O}/^{16}\text{O}$) в спирте. При исследовании изотопии легких элементов с помощью МСИО невозможно установить различия между этанолом из углеводов любого представителя группы растений СЗ-пути фотосинтеза (сахарной свеклы, яблок, пшеницы, картофеля и др.), в которую входит и виноград, если они произрастают в одном агроклиматическом регионе. Причина заключается в том, что все они обладают единым механизмом изотопного обмена в метаболических процессах.

В то же время последние разработки в этой области позволяют проводить селективные измерения общего состава изотопов водорода $^2\text{H}/^1\text{H}$ в метильной и метиленовой группах молекулы этанола после его дегидратации (метод EIM-IRMS/SIRA).

МЕТОДИЧЕСКИЕ ОСНОВЫ КОМПЛЕКСНОГО СКРИНИНГА ЯМР $^2\text{H}/^1\text{H}$ (CUMULATIVE SCREENING NMR/CS-NMR)

Спектроскопия ЯМР минорных изотопов ^2H , ^{13}C , ^{15}N , ^{17}O и ^{33}S обеспечивает в общем случае измерение их дифференциального распределения во фрагментах молекул. На практике этот способ можно использовать только для водорода, у которого диапазон вариации природного содержания значителен (около 150 млн $^{-1}$). В природных водах он составляет почти 100 млн $^{-1}$, а его дифференциация даже в отдельных фрагментах органических молекул (например, α -пинена) достигает 150 млн $^{-1}$. Неравновесное распределение изотопа водорода – дейтерия ($^2\text{H}(\text{D})$) в органических молекулах является изотопным паспортом, подделка которого практически невозможна. Однако этот метод трудно

Рис.1. Спектр ЯМР ^1H образца вина в условиях устранения сигналов от сателлитов ^{13}C : 1 – глицерин (10,01 г/л), 2 – метанол (153,00 мг/л), 3 – янтарная кислота (868,67 мг/л), 4 – уксусная кислота (512,72 мг/л), 5 – молочная кислота (828,28 мг/л), 6 – 2,3-бутандиол (414,33 мг/л)

применять для изотопов других элементов из-за малой вариабельности их содержания в природных объектах по отношению к прецизионности измерений интегральных интенсивностей сигналов ЯМР.

Все ВОР объединяет главная составляющая – вода. Поэтому сначала рассмотрим вопрос определения содержания в ней дейтерия $^2\text{H}(\text{D})$ методом спектроскопии ЯМР. Если объект анализа – природная вода с незначительным содержанием органических и неорганических примесей (до 3,5% в среднеокеанической воде), можно использовать традиционный подход МСИО, в котором водород воды для измерения переводится в газовой фазе в молекулярный при нагревании до 400° в присутствии цинка [1] или иным каталитическим путем [2]. Спектроскопия ЯМР является прямым способом измерения абсолютного содержания дейтерия в воде путем количественного введения в раствор добавок калибровочного соединения с известным содержанием дейтерия (например, тетраметилмочевины или диметилсульфона). Поскольку чувствительность метода ЯМР ^2H на шесть порядков ниже, чем для основного изотопа водорода – протия ^1H , то при исследованиях необходимо минимизировать потери объема анализируемой воды, сокращая долю калибровочных соединений. Для этого мы предлагаем использовать специальные обогащенные на 1–2 порядка дейтерорастворители для ЯМР, что позволяет уменьшить их объемную долю в растворе аналита до 2–3%. В работе [3] в качестве калибровочного соединения применяли триметилсилилпропионат

натрия-2,2,3,3- d_4 (ТСП), сигналы дейтерия двух метиловых групп которого имеют химические сдвиги 0,60 и 2,00 м.д. Такой внутренний калибратор при измерении природного содержания дейтерия в воде неудобен из-за необходимости многократно понижать его концентрацию в образце, например, разбавляя в 10^4 – 10^5 раз уксусной кислотой, сигнал гидроксильной группы которой также следует учитывать. В работе [4] предложен иной способ калибровки – с помощью внешнего стандарта. Он содержит обогащенную дейтерием воду и реагент трифторметансульфонат европия ($\text{Eu}[\text{CF}_3\text{SO}_3]_3$), смещающий сигнал дейтерия воды внешнего стандарта в слабое поле на 2–3 м.д., что исключает его перекрывание с пиком исследуемого образца. Однако работа с внешним стандартом имеет ряд недостатков. Для ампул ЯМР диаметром 5 мм за счет потери объема аналита на 15–20% время каждого измерения увеличивается на 30–40% в зависимости от содержания воды в ВОР. Необходимы тщательно откалиброванные по объему пары ампул (стандартной и капилляра), поэтому усложняется проведение серий экспериментов, особенно с использованием устройств автоматической смены образцов.

Для определения содержания воды в ВОР необходимо знать ее объемную долю, что требует идентификации и измерения остальных компонентов на основе устанавливаемой из ЯМР ^1H их мольной доли. В расчете следует учитывать вклад в интегральную интенсивность сигнала ^1H воды протонов других компонентов, участвующих с ней в быстром обмене. Без учета этих факторов определение по площади сигнала в области 4,8 м.д. всегда дает заниженное значение. Идентификация и количественное определение из спектра ЯМР ^1H сокомпонент в сложных системах позволяют рассчитать содержание непосредственно воды, а из спектра ЯМР ^2H установить ее изотопный состав, важный для исследовательских целей или характеристики подлинности.

КОМПЛЕКСНЫЙ ЯМР-СКРИНИНГ ВОДНОЙ КОМПОНЕНТЫ ВИНОДЕЛЬЧЕСКОЙ ПРОДУКЦИИ

Рассмотрим процедуру определения в ВОР содержания дейтерия воды на примере экспертной задачи быстрого скрининга подлинности часто фальсифицируемых натуральных вин. Под скринингом подразумевается проведение оперативного количественного анализа состава объекта простыми методами без пробоподготовки с возможностью немедленного принятия решения [5].

Специфика различий компонентного и изотопного состава вин и другой алкогольной продукции (напри-

мер, коньяков, коньячных спиртов) обусловлена агро-климатическими и технологическими особенностями региона культивирования и переработки винограда, в том числе количеством осадков, температурой, интенсивностью ирригационных мероприятий, составом поверхностных и подземных вод, почвой и др. Эти факторы влияют, в частности, на содержание в вине минорных компонентов и изотопный состав винограда.

При анализе вина первоначально регистрируется спектр ЯМР ^1H , из которого с необходимой для идентификационных целей достоверностью легко рассчитать содержание воды, этанола, глицерина, основных кислот – винной, яблочной, молочной, лимонной, янтарной, уксусной.

В качестве маркера при измерении содержания определенных минорных компонентов удобно использовать сильнополюсный сателлит группы CH_3 этанола, который обусловлен присутствием молекул, в состав которых входит изотоп углерода ^{13}C (1,1%). Количественное определение других минорных компонентов следует выполнять в режиме устранения сателлитных сигналов, то есть регистрируя спектры ЯМР ^1H при облучении образца на частоте ядер ^{13}C . В качестве примера на рис.1 представлен спектр ЯМР ^1H вина "Каберне Совиньон" урожая 2015 года (Анапа), в котором содержание этанола рассчитано по интегральным интенсивностям сигналов групп OH , CH_3 и CH_2 (на рис.1 они частично обрезаны) и составляет 13,9 об. %. Видны учитываемые при расчете сигналы многих минорных компонентов, которые характеризуют подлинность вина с позиций компонентного состава.

Хотя спектры ЯМР ^1H прямо характеризуют содержание отдельных компонентов, их недостаточно для признания подлинности вина или его соответствия определенным агроклиматическим условиям выращивания винограда (географическому происхождению). Идентичный представленному на рис.1 спектр ЯМР ^1H можно имитировать искусственной смесью воды, этанола и минорных компонентов. Сигнал воды в спектрах ЯМР ^1H и ^2H имеет характеристическое положение (4,8 м.д.) и не перекрывается с пиками других компонентов. Его интегральную интенсивность в спектре ЯМР ^2H следует использовать для аутентификации вина при учете вклада водорода гидроксильных и карбоксильных групп остальных ингредиентов, в первую очередь этанола. Для строгого заключения желательно знать содержание дейтерия в воде подлинных вин этого региона и агроклиматические условия заявленного года сбора урожая. Более доступный критерий – сравнение содержания дейтерия в воде вина и его усредненного значения в атмосферных осадках и почвенных водах

региона. Известно, что количество дейтерия в воде вина на 8–15 млн⁻¹ выше, чем в осадках из-за особенностей изотопного фракционирования в молекулах углеводов (сахаров) винограда при их превращении в этанол в ходе брожения. На территории винодельческих регионов РФ все поверхностные и подземные воды содержат менее 150 млн⁻¹ дейтерия, поэтому в водной фракции вина его должно быть больше 156–158 млн⁻¹. Следовательно, представленный метод ЯМР способен выявлять разбавление вина смесью этанол/вода или водой, а также полную фальсификацию его состава экзогенными компонентами.

Отметим, что представленный метод ЯМР при исследовании изотопии водорода в водной фракции продукта не позволяет обнаружить добавление в сусло экзогенных сахаров других растений C3- и/или C4-пути фотосинтеза (напр., сахарной свеклы, сахарного тростника, кукурузы, сорго и др.) перед брожением (т.н. "шапталлизация") или непосредственное введение в винодельческую продукцию спирта невиноградного происхождения. Для этого необходимо измерить содержание дейтерия в метильной (показатель (D/H)_{II}) и метиленовой (показатель (D/H)_{III}) группах молекулы этанола, выделенного из исследуемого вина. Поэтому измерение содержания дейтерия в воде вина является первым быстрым этапом его детального анализа, обеспечивающего выявление наиболее грубых фальсификаций.

Представленный алгоритм, включающий комплексный анализ спектров ЯМР протия – ¹H и дейтерия – ²H (D) вина для измерения содержания в его воде изотопа ²H впервые был предложен нами для дифференциации вин Грузии, России (регион Кубани) и стран Западной Европы 15 лет назад [6]. Основные результаты этого исследования отражает рис.2, на котором за нулевое значение принято содержание дейтерия в международном стандарте VSMOW, соответствующем экваториальной воде мирового океана (D/H = 155,76 млн⁻¹). Количество дейтерия в водной фракции вин Грузии, Испании, Италии, Франции близко, но выше, чем в винах Краснодарского края и Ростовской области. Частично это обусловлено агроклиматическими причинами и, казалось бы, может быть выбрано в качестве критерия при дифференциации вин Грузии и России. Однако в последних содержание дейтерия в водной фракции варьирует в широких пределах. Возможные причины: влияние агроклиматических условий выращивания винограда и интенсивности ирригационных мероприятий, а также добавление в некоторые образцы воды или смеси вода/этанол в типичном для вин соотношении. Из спектров ЯМР ¹H выявлены манипуляции с со-

Рис.2. Относительное содержание дейтерия в водной компоненте вин различных географических регионов [6]. За нулевое значение принято содержание дейтерия в стандарте среднеокеанической воды VSMOW, для которого D/H = 155,76 млн⁻¹

ставом некоторых вин по соотношению в них этанола, глицерина и кислот. Более того, в трех образцах содержание дейтерия в водной фракции вина оказалось ниже значений для поверхностных вод этого региона (145–150 млн⁻¹), что нереально. Такой подход к выявлению фальсификации натуральных вин в последующие годы не применялся, как это следует из обзорной статьи 2015 года [7].

Разработанный нами алгоритм анализа прошел апробацию в 2017 году, когда из конкурсных вин выставки VINORUS, проходившей в Краснодаре совместно с 1-й Международной научно-технической конференцией "Инновационный мир современного виноградарства и виноделия: Россия – InnoWine Russia-2017", были отобраны образцы красного сухого вина сортов винограда "Каберне Совиньон" и "Мерло" 2015 года из различных регионов: Краснодарского края (№ 2–11), Крыма – Бахчисарай (№ 12), Волгоградской области (№ 13), Испании (№ 14). По спектрам ЯМР ¹H образцов измерены содержания в них этанола и глицерина (табл.1). Вклад в сигналы ¹H и ²H воды гидроксильных и карбоксильных групп глицерина и других минорных компонентов вина не учитывали. В табл.1 приведены измеренные методом ЯМР относительные количества дейтерия в водной фракции различных вин. Для наглядности они представлены на рис.3 вместе с данными по содержанию ²H в водопроводной воде Краснодара от 28 апреля 2017 года (№ 1). В водной фракции вина

Таблица 1. Показатели вина, полученные из спектров ЯМР

№ образца	Доля водной фракции в вине, об. %	Этанол, об. %	Глицерин, г/л	Относительное содержание дейтерия в водной фракции вина, млн ⁻¹
2	87,17	12,83	6,67	149,1
3	86,91	13,09	5,37	166,4
4	86,64	13,36	8,60	165,6
5	87,42	12,58	5,41	165,6
6	87,17	12,83	5,76	151,7
7	85,76	14,24	8,53	166,3
8	87,06	12,94	7,85	166,2
9	85,30	14,70	8,44	162,4
10	86,58	13,42	8,12	159,1
11	88,32	11,68	5,26	164,2
12	86,06	13,94	8,64	163,8
13	84,75	15,25	10,01	161,9
14	86,70	13,30	8,78	167,6

всех исследованных проб содержание дейтерия ²H варьирует в диапазоне 159–167 млн⁻¹. Исключение составили два образца вина "Каберне Совиньон" региона Тамани (№ 2 и 6) с пониженным содержанием дейтерия в воде, что можно объяснить составом воды источника, который использовали для орошения соответствующего винограда.

ЯМР-СПЕКТРОСКОПИЯ ДЕЙТЕРИЯ В ЭТАНОЛЕ ВИНОДЕЛЬЧЕСКОЙ ПРОДУКЦИИ

Наиболее широкую известность и прикладное значение метод ЯМР ²H получил в исследовании содержания дейтерия в метильной (D/H)_I и метиленовой (D/H)_{II} группах молекулы этанола для установления его природы.

Характеризация ВОР по содержанию дейтерия во фрагментах молекул этанола вина требует его количественного выделения. Для этого детально разработана соответствующая процедура, так как в различных странах разрешено добавление в виноградное сусло экзогенных сахаров, в том числе из растений других ботанических видов.

В 1990 году Международной межправительственной организацией по виноградарству и виноделию OIV в Евросоюзе в качестве официального принят впервые описанный в [8] метод SNIF-NMR [9]. К сегодняшнему дню накоплена значительная база научных знаний об интервалах изменения содержания дейтерия в метильной (D/H)_I и метиленовой (D/H)_{II} группах молекулы этанола виноградного проис-

хождения. С учетом агроклиматических (географических) и технологических факторов минимальный и максимальный уровни для показателя (D/H)_I в виноградном этаноле составляют 98,0 и 106,0 млн⁻¹ соответственно, а для показателя (D/H)_{II} – 121,0 и 136,0 млн⁻¹. В эти диапазоны попадают также этанолы, выделенные из некоторых других фруктов, но не из сахарной свеклы (как и виноград входит в группу растений C3-пути фотосинтеза) и растений C4-пути фотосинтеза (кукуруза, сахарный тростник, сорго и др.) [10–11].

Метод SNIF-NMR требует использования спектрометров ЯМР, с рабочей частотой не менее 400 МГц для ядер ¹H и оснащенных датчиком для ампул 10 мм и системой стабилизации по сигналу ЯМР ¹⁹F.

В Центре коллективного пользования (Научно-образовательном центре) РУДН для решения задачи по исследованию природы этанола в продуктах виноделия используется модифицированный метод SNIF-NMR, основанный на авторской разработке [12] для стандартных спектрометров ЯМР.

Одна из его особенностей состоит в том, что стабилизация условий резонанса по сигналу ¹⁹F не применяется, поэтому за счет небольшого дрейфа сигналов измерение содержания дейтерия в CH₂- (D/H)_{II} и CH₃- (D/H)_I фрагментах молекул этанола по высоте соответствующих пиков имеет большие погрешности. Однако это практически не влияет на площади, измеряемые, как это принято в количествен-

Рис.3. Относительное содержание дейтерия в водной компоненте образцов красного вина урожая 2015 года. За нулевое значение принято содержание дейтерия в стандарте среднеокеанической воды VSMOW, для которого $D/H = 155,76 \text{ млн}^{-1}$

ной спектроскопии ЯМР, по интегральной, а не пиковой интенсивности. Несомненное преимущество такого подхода перед методом [9] состоит в том, что содержание дейтерия в гидроксильной группе ОН выделенного этанола или остаточной воды можно измерять по интенсивности сигналов, которые за счет обменных процессов всегда имеют большую ширину, чем у дейтерия ^2H в метильной CH_3 (D/H)_I и метиленовой CH_2 (D/H)_{II} групп. Поскольку содержание дейтерия в гидроксильной группе определяется исключительно водной фракцией вина, в детальном анализе спектра ЯМР ^1H необходимости нет. Достаточно измерить плотность раствора анализата для оценки содержания в этаноле остаточной воды. Другая особенность разработанного метода при измерении природного содержания изотопа ^2H – применение в качестве калибровочной добавки обезвоженного диметилсульфоксида (ДМСО) с повышенным на два порядка содержанием дейтерия, что в 10 раз (с 30 об. % в случае метода [9] до 3 об. %) уменьшает объемную долю калибровочного вещества и, тем самым, вдвое сокращает время регистрации ЯМР-спектра. Суммарная погрешность такого способа обеспечивает прецизионность измеряемых значений $\pm 3\%$, что вполне достаточно для оценки природы этанола.

Разработку метода и его апробацию выполняли на спектрометре JEOL JNM-ECA 600 (Япония) с рабочей частотой для ядер ^1H 600 МГц и для ядер ^2H 92,102 МГц. Прибор снабжен системой автоматической замены образцов и имеет датчики 5 и 10 мм (рис.4). В работе использовали откалиброванные ампулы диаметром $(4,97 \pm 0,013) \text{ мм}$ и длиной 178 мм. В качестве

Рис.4. ЯМР спектрометр JEOL JNM-ECA 600 (Япония) с рабочей частотой для ядер ^1H 600 МГц и для ядер ^2H 92,102 МГц. Сотрудник ЦКП (НОЦ) РУДН – В.Ивлев

внутреннего стандарта для количественного определения компонентов при регистрации спектров ЯМР ^1H использовали триметилсилилпропионат натрия- d_4 (ТСП), отклик метильных протонов которого не перекрывается с сигналами анализируемых образцов. При регистрации спектров ЯМР ^2H в качестве внутреннего стандарта (3% об.) использовали ДМСО с искусственно повышенным на два порядка содержанием дейтерия. Типичный спектр представлен на рис.5. Для регистрации спектров на ядрах ^2H выбраны следующие условия эксперимента: импульс 45° , усиление 72, смещение 5 м.д., развертка 11 м.д., 16К точек на спектр, время наблюдения спада свободной индукции 2 с, число сканирований 800, время задержки между импульсами $5 \cdot T_1(^2\text{D}_2\text{O}) \geq 5 \text{ с}$. Для обработки спектра использовалась автоматическая коррекция базовой линии, ручная настройка фазы, экспоненциальное умножение на значение 2,0.

В табл.1 представлены усредненные данные пяти измерений содержания дейтерия ^2H в каждом образце, относительное стандартное отклонение для которых составляет 2%.

ЗАКЛЮЧЕНИЕ

Предложен новый метод совокупного ЯМР-скрининг-анализа (CS-NMR) для определения содержания дейтерия в воде различных водно-органических растворов, в том числе вина, винодельческой продукции, соков и т.п. Первый этап – регистрация количественных спектров ЯМР ^1H для определения содержания воды в образце, с учетом других идентифицированных компонентов раствора. Затем – полу-

Рис.5. Спектр ЯМР ^2H образца спиртового дистиллята с добавлением ДМСО: 1 – ОН (гидроксильная группа этанола), 2 – H_2O (остаточная вода в дистилляте), 3 – CH_2 (метиленовая группа этанола ($\text{D}/\text{H}_{\text{II}}$), 4 – калибровочная добавка ДМСО, 5 – CH_3 (метильная группа этанола ($\text{D}/\text{H}_{\text{I}}$))

чение спектра ЯМР ^2H , из которого рассчитывается содержание изотопа ^2H воды, как важной характеристики происхождения или аутентичности с контрольным образцом. Не требующий пробоподготовки метод удобен при быстром скрининге многих водно-органических объектов. Детальное определение изотопного состава водорода других компонентов ВОР после их выделения может выполняться на стандартном оборудовании по более быстрой оригинальной методике.

Для исследования природы этанола винодельческой и другой спиртосодержащей продукции в Центре коллективного пользования (НОЦ) РУДН применяется модифицированный метод ЯМР, который основан на определении содержания дейтерия в метильной (показатель ($\text{D}/\text{H}_{\text{I}}$)) и метиленовой (показатель ($\text{D}/\text{H}_{\text{II}}$)) группах молекулы спирта. Количественные данные о содержании дейтерия в метильной группе ($\text{D}/\text{H}_{\text{I}}$) спирта позволяют сделать вывод не только о ботаническом виде растения, но и о географическом регионе его происхождения, а также о природе воды, участвовавшей в процессе брожения углеводов с образованием этанола. Данные о дейтерии в метиленовой группе ($\text{D}/\text{H}_{\text{II}}$) спирта свидетельствуют о климатических условиях произрастания растения (например, винограда).

ЛИТЕРАТУРА

1. Hachey D.I., Wong W.W., Boutton T.W., Klein P.D. Isotope ratio measurements in nutrition and biomedical

research // Mass Spectrometry Reviews. 1987. V. 6. P. 289–328. DOI 10.1002/mas.1280060203

2. Scrimgeour C.M., Rollo M.M., Mudambo S.M., Handley L.I., Prosser S.J. A simplified method for deuterium/hydrogen isotope ratio measurements on water samples of biological origin // Biological Mass Spectrometry/ 1993. 22. P. 383–387. DOI 10.1002/bms.1200220704
3. Ma X., Deng P., Wang X., Zhang C., Hou X. Direct determination of deuterium of wide concentration range in water by Nuclear Magnetic Resonance // Talanta. 2012. 97. P. 450–455.
4. Джимаков С.С., Басов А.А., Федюлова Л.В., Быков И.М., Ивлев В.А., Мелконян К.И., Тимаков А.А. Определение концентрации дейтерия в биологических жидкостях с помощью ЯМР-спектроскопии // Авиакосмическая и экологическая медицина. 2016. 50(3). С. 42–47.
5. Varcarel M., Caerdenas S., Gallego M. Sample screening systems in analytical chemistry // Trends in Analytical Chemistry. 1999. 18(11). P. 685–694.
6. Калабин Г.А., Козлов Ю.П., Рыков Р.С., Кулагина Н.В., Кушнарев Д.Ф., Рохин А.В. Новый алгоритм идентификации натуральных вин Грузии методом спектроскопии ЯМР // Вестник РУДН. Серия "Экология и безопасность жизнедеятельности". 2003. 9. С. 77–82.
7. Christoph N., Hermann A., Wachter H. 25 Years authentication of wine with stable isotope analysis in the European Union – Review and outlook // BIO Web of Conferences. 2015. 5. 02020. 8 p.
8. Martin G.J., Martin M.L., Mabon F., Michon M.J. Identification of the origin of natural alcohols by natural abundance hydrogen-2 nuclear magnetic resonance // Analytical Chemistry. 1982. 54. P. 2380–2382.
9. EC-Regulation 2676/1990 // Official Journal of the European Community. 1990. 64. L272.
10. Christoph C.B., Wachter H., Christoph N., Rossmann A., Adam L. Assignment of raw material and authentication of spirits by gas chromatography, hydrogen- and carbon-isotope ratio measurements I. Analytical methods and results of a study of commercial products // Zeitschrift fuer Lebensmitteluntersuchung und – forschung A. 1997. 204. P. 445–452.
11. Perini M., Guzzon R., Simoni M., Malacarne M., Larcher R., Camin F. The effect of stopping alcoholic fermentation on the variability of H, C and O stable isotope ratios of ethanol // Food Control. 2014. 40. P. 368–373.
12. Калабин Г.А., Кулагина Н.В., Рыков Р.С., Козлов Ю.П., Кушнарев Д.Ф., Рохин А.В. Идентификация сырьевой природы этанола методом спектроскопии ЯМР // Вестник РУДН. Серия "Экология и безопасность жизнедеятельности". 2003. 7. С. 87–98.

DETERMINATION OF DEUTERIUM CONTENT IN COMPONENTS OF WATER-ORGANIC SOLUTIONS

BY A COMBINATION OF THE NMR ^1H AND ^2H METHODS*

G.A.Kalabin, V.A.Ivlev, N.A.Komarov, A.Yu.Kolesnov,
Peoples' Friendship University of Russia (RUDN University)
kalabinga@mail.ru, irms@biolab.ru

An algorithm is proposed for using the combination of NMR spectroscopy of ^1H and deuterium $^2\text{H(D)}$ nuclei for prompt identification and determination of the component and differential isotope composition of water organic solutions including $(\text{D/H})_I$ and $(\text{D/H})_{II}$ values in ethanol, water in wine, cognac, brandies, juices, soft drinks and other products. The method does not require sample preparation and/or use of any certified reference materials. Some examples of the proposed concept application for the authentication and quality control of wine products are presented.

INTRODUCTION

The composition of any water-organic solutions (WOS), for example, alcoholic beverages, milk, wines, juices etc. containing H, C, N, O and S elements can be characterised at various hierarchical levels, such as component, fragment, atomic and isotopic ones. A comprehensive analysis of the obtained complementary data makes it possible to identify the components and origin of WOS with a high degree of accuracy. The elemental and component analysis methods were the most widely used for this purpose. The latter predominantly uses different types of chromatography, which, in combination with the mass spectrometric detection, is highly sensitive and makes it possible to identify and quantify the content of macro- and microcomponents in the presence of certified reference samples (CRS). In studying the WOS of an unknown composition, the quantitative nuclear magnetic resonance (NMR) ^1H and ^{13}C spectroscopy, which does not require the use of CRS, offers some real advantages. However, even a combination of these methods does not provide a rigorous proof of the conformity of the WOS components to the declared varietal, agroclimatic (geographical) and/or technological origin, for example, wines- protected geographical indication (PGI) or the protected

designation of origin (PDO), mineral water – source area, juices - NFC or reconstituted, milk – whole, adulterated or reconstituted milk. The WOS component composition, if economically feasible, can be simulated by preparing a solution of pure water as well as all necessary and widely available inorganic and organic components. However, a counterfeit, even if the formulation components are carefully manufactured and selected, will not correspond to the genuine product in terms of the content of minor stable isotopes of hydrogen for example (^2H), carbon (^{13}C), nitrogen (^{15}N), oxygen (^{17}O & ^{18}O) and sulphur (^{33}S & ^{34}S). It is only isotopic analysis that can make it possible to discover a counterfeit.

The isotope ratio mass spectrometry (IRMS/SIRA) takes the lead in the isotopic analysis due to the highest sensitivity and precision. The initial difficulties in the application of this method were associated with the most difficult stage, the isolation of components of an object and their purification for subsequent authentication. The IRMS toolkit development, for example, the introduction of Gas Chromatography (GC) (for the evaluation of volatile compounds) and High Performance Liquid Chromatography (HPLC) systems (for the evaluation of non-volatile compounds) combined with the corresponding isotope interfaces made it

* The publication has been prepared with the support of the RUDN University Program 5–100.

possible to exclude the difficult and time consuming sample preparation associated with extraction and purification of analytes from the matrix. As a rule, the so-called integral ratios of the isotopes $^2\text{H}/^1\text{H}$, $^{13}\text{C}/^{12}\text{C}$, $^{15}\text{N}/^{14}\text{N}$, $^{18}\text{O}/^{16}\text{O}$, $^{34}\text{S}/^{32}\text{S}$ are determined by the IRMS/SIRA method. For example, for the ethyl alcohol molecule $\text{CH}_3\text{CH}_2\text{OH}$, only values of $^2\text{H}/^1\text{H}$ ($\delta^2\text{H}(\text{D})$), $^{18}\text{O}/^{16}\text{O}$ ($\delta^{18}\text{O}$) or $^{13}\text{C}/^{12}\text{C}$ ($\delta^{13}\text{C}$) can be measured in the composition of wine, although these isotopes can be characterised by a non-equilibrium distribution in individual fragments of molecules which is typical for agroclimatic conditions of the geographical region of grapes cultivation, i.e. three and two different values for hydrogen and carbon respectively. In addition, when studying, for example, alcohol-containing distillates, the presence of water, which influences the reliability of the results when measuring isotope ratios of hydrogen ($^2\text{H}(\text{D})/^1\text{H}$) and oxygen ($^{18}\text{O}/^{16}\text{O}$) in alcohol, should be taken into account.

When studying the isotopy of light elements by the IRMS/SIRA method only, and in the case of coincidence of agroclimatic (geographical) conditions of plant growth, it is impossible to measure differences between ethanol from carbohydrates of any representative from the C3-pathway of photosynthesis (e.g. grapes, sugar beets, apples, wheat, potatoes, etc.) due to a single mechanism of isotope metabolism in this group of plants. At the same time, the latest developments IRMS/SIRA mass-spectrometry allow selective measurements of the total composition of hydrogen isotopes $^2\text{H}/^1\text{H}$ in the methyl and methylene groups of the ethanol molecule after its dehydration (EIM-IRMS/SIRA method).

METHODOLOGICAL BASIS FOR CUMULATIVE NMR SCREENING $^2\text{H}/^1\text{H}$ (CUMULATIVE SCREENING NMR /CS-NMR)

In the general case, the NMR spectroscopy of the minor isotopes ^2H , ^{13}C , ^{15}N , ^{17}O and ^{33}S provides the measurement of their differential distribution in molecule fragments. In practice, this method can be used only for hydrogen which has a significant range of variation of the natural content (about 150 ppm). In natural waters, it is almost 100 ppm, and its differentiation even in individual fragments of organic molecules (for example, α -pinene) reaches 150 ppm. The non-equilibrium distribution of the hydrogen isotope deuterium ($^2\text{H}(\text{D})$) in organic molecules is an isotope passport which is practically impossible to falsify. However, this method is difficult to use for the isotopes of other elements due to the low variability of their content in natural objects with respect to the precision of measurements of the integrated intensities of NMR signals.

Fig.1. The ^1H NMR spectrum of the wine sample under the conditions of elimination of signals from ^{13}C : 1 – glycerin (10.01 g/l), 2 – methanol (153.00 mg/l), 3 – succinic acid (868.67 mg/l), 4 – acetic acid (512.72 mg/l), 5 – lactic acid (828.28 mg/l), 6 – 2,3-butanediol (414.33 mg/l)

All WOS have the main component in common, water. Therefore, let us first consider the determination of the content of deuterium $^2\text{H}(\text{D})$ in it by NMR spectroscopy. If the analysis object is natural water with an insignificant content of organic and inorganic impurities (up to 3.5% in mid-ocean water), a traditional IRMS/SIRA approach can be used, in which the water hydrogen for measurement is transferred in the gas phase to the molecular one when heated to 400° in the presence of zinc [1] or in another catalytic way [2]. NMR spectroscopy is a direct method for measuring of the absolute content of deuterium in water by quantitative addition of a calibrated compound with the known deuterium content (e.g. tetramethylurea or dimethylsulfoxide) to the solution. Since the sensitivity of the ^2H NMR method is by six orders of magnitude lower than that for the main isotope of hydrogen – protium ^1H , during the measurement it is necessary to minimise the loss of volume of the water being analysed by reducing the proportion of calibration compounds. To this end, we propose to use special calibration compounds enriched of ^2H by 1–2 orders of magnitude for NMR, which makes it possible to reduce their volume fraction in the analyte solution to 2–3%. In [3], sodium trimethylsilylpropionate-2,2-3,3- d_4 (TSP) was used as the calibration compound, the deuterium signals of two methylene groups of which have chemical shifts of 0.60 and 2.00 ppm. In measuring the natural content of deuterium in water, such an internal calibrant is inconvenient because of the need to somehow reduce its concentration by

several times in the sample, for example, by diluting by 10^4 – 10^5 times with acetic acid, the hydroxyl group signal of which should also be taken into account. In [4] another calibration method was proposed, with an external standard. It contains deuterium-enriched water and europium trifluoromethanesulfonate reagent ($\text{Eu}[\text{CF}_3\text{SO}_3]_3$), which shifts the deuterium signal of the external standard water into a weak field by 2–3 ppm, which excludes its overlapping with the peak of the test sample. However, working with an external standard has a number of drawbacks. For 5 mm diameter NMR tubes, due to the loss of analyte volume by 15–20%, the time of each measurement increases by 30–40% depending on the water content in the WOS. The pairs of tubes (standard and capillary) carefully calibrated by volume are necessary; therefore, it is more difficult to conduct a series of experiments, especially with automatically changing samples.

To determine the water content in the WOS, it is necessary to know its volume quantity, which requires the identification and measurement of the remaining components based on their molar fraction established from NMR ^1H measurements. In the calculation it is necessary to take into account the contribution to the integral intensity of the protons of other components engaged in the rapid exchange with it of the ^1H signal. Without taking these factors into account, the signal area determination in the region of 4.8 ppm. always gives an understated value. Identification and quantitative determination from the NMR ^1H spectrum of the co-component in complex systems allow the calculation of the water content directly; and from the NMR ^2H spectrum, establish its isotopic composition important for research purposes or the authenticity characterisation.

INTEGRATED NMR SCREENING OF THE WATER COMPONENT OF WINES AND SIMILAR PRODUCTS

Let us consider the procedure for determining the content of deuterium water in the WOS by using the example of the expert task concerning quickly screening the authenticity of often adulterated wines. Screening means the operative quantitative analysis of the composition of the object by simple methods without any sample preparation with the possibility of immediately taking a decision [5].

The specificity of differences in the component and isotope composition of wines and other alcohol products (for example, cognacs, brandies or spirits) is due to the agroclimatic and technological peculiarities of the region of cultivation and processing of grapes including

atmospheric precipitation, temperature, intensity of irrigation measures, surface and groundwater composition, soil etc. In particular, all these factors influence the content of minor components in wine and isotopic composition of grapes.

In the analysis of wine, the NMR ^1H spectrum is initially recorded, from which it is easy to calculate the content of water, ethanol (vol. %), glycerol, basic acids such as tartaric, malic, lactic, citric, succinic and acetic as well as identify residual carbohydrates, if any.

In measuring the content of certain minor components, it is convenient to use the high-field satellite of the CH_3 ethanol group as a marker which is due to the presence of molecules containing the carbon ^{13}C isotope (1.1%). The quantification of other minor components should be performed in the mode of satellite signal elimination, i.e. by recording the NMR ^1H spectra in the sample radiation at a frequency of nuclei ^{13}C . Fig.1 shows as an example the NMR spectrum of the ^1H Cabernet Sauvignon wine of the 2015 crop (Anapa), in which the ethanol content (13.9 vol. %) is calculated from the integrated intensities of the signals of the OH, CH_3 and CH_2 groups (they are partially cut off in Fig.1). Signals taken into account during the calculation of many minor components, which characterise the wine authenticity from the viewpoint of the component composition, can be seen.

Although NMR ^1H spectra directly characterise the content of individual components, they are insufficient to recognize the wine authenticity or its compliance with certain agroclimatic conditions of grapes cultivation (geographic origin). A spectrum identical to the ^1H NMR shown in Fig. 1 can be simulated by an artificial mixture of water, ethanol and minor components. The water signal in the NMR ^1H & ^2H spectra has a characteristic position (4.8 ppm), nor it overlaps with peaks of other components. Its integral intensity in the NMR ^2H spectrum should be used for wine authentication when taking into account the hydrogen contribution of the hydroxyl and carboxyl groups of the remaining ingredients, primarily ethanol. For a strict expert opinion, it is desirable to know the deuterium content in water of the authentic wines of this region and the agroclimatic conditions of the declared harvest year. A more accessible criterion is a comparison of deuterium content in wine water and its averaged value in precipitation and soil waters of the region. It is known that the amount of deuterium in the wine water is 8–15 ppm higher than in sediments because of the peculiarities of isotope fractionation in the molecules of carbohydrates (sugars) of grapes when they are converted to ethanol during

fermentation. On the territory of the Russian wine-growing regions, all surface and groundwater contain less than 150 ppm deuterium, so in the water fraction of the wine it should be more than 156–158 ppm. Therefore, the presented NMR method is able to detect the dilution of wine with a mixture of ethanol/water or water as well as a complete falsification of its composition with exogenous components. At the same time, it should be that by studying the isotopy of hydrogen in the water fraction of the product on the basis of the NMR method, it is not possible to detect the addition to musts of exogenous sugars of other plants of the C3 and/or C4 photosynthesis pathways (e.g. sugar beet, sugar cane, maize, sorghum and others) before fermentation ("shaptalization") or direct introduction into wines of alcohol of the non-grape origin. To do this, it is necessary to measure the content of deuterium in the methyl ((D/H)_i) and methylene ((D/H)_{ii}) groups of the ethanol molecule isolated from the wine. Therefore, the measuring the content of deuterium in wine water is the first rapid stage of its detailed analysis to ensuring that the most gross counterfeits can be detected.

The algorithm including an integrated analysis of the NMR spectra of protium ¹H and deuterium ²H of wine, was first proposed by us for differentiating the wines of Georgia, Russia (the Kuban region) and Western European countries 15 years ago [6]. The main results of the research are shown in Fig.2, where the value 0 is the deuterium content in the international VSMOW standard corresponding to the equatorial water of the world ocean (155.8 ppm). The amount of deuterium in the water fraction of wines in Georgia, Spain, Italy and France is close, but higher than in the wines of the Krasnodar and the Rostov Regions. This is partly due to agroclimatic reasons and seems to be chosen as a criterion in the differentiation of the wines of Georgia and Russia. However, in the latter, the content of deuterium in the water fraction varies widely. Possible causes include the influence of agroclimatic conditions of grape growing and the intensity of irrigation measures as well as the addition of some water or a water/ethanol mixture to the level typical for wines. Manipulations with the composition of some wines were revealed from NMR ¹H spectra by the ratio of the content of ethanol, glycerin and acids in them. Moreover, in three samples, the content of deuterium in the water fraction of the wine turned out to be lower than the values for the surface waters of this region (145–150 ppm), which is unrealistic. Such an approach to the identification of counterfeit of natural wines in the following years was not applied, as follows from the review of the article of 2015 [7].

Fig.2. The content of deuterium in the water fraction of wines of different geographical regions [6]. Samples are grouped by country of wine production

The analysis algorithm developed by us was tested in 2017, when wine samples were selected at the VINORUS trade exhibition held in Krasnodar along with the 1st International Scientific and Technical Conference Innovative World of Modern Viticulture & and Wine-Making: Russia (InnoWine: Russia-2017). There were samples of dry wine of Cabernet Sauvignon and Merlot grapes (vintage 2015) from different regions - Krasnodar region (No 2–11), Crimean peninsula - Bakhchisaray (No 12), Volgograd region (No 13) and Spain (No 14). The contents of ethanol and glycerol were measured by the NMR ¹H spectra analysis of samples (Table 1). The contribution to the signals ¹H and ²H of the water of the hydroxyl and carboxyl groups of glycerin and other minor wine components was not taken into account. Table 1 shows the relative amounts of deuterium in the water fraction of various wines measured by the NMR method. For clarity, they are presented in Fig.3, together with data by the content of ²H in the tap water of Krasnodar dated 28 April 2017 (No 1). In the water fraction all of tested wine, the deuterium ²H content varies in the range from 159 to 167 ppm. Two samples of Cabernet Sauvignon wine from the Taman area of the Krasnodar region (No 2 and 6) were found as an exception because of the reduced deuterium content in water which can be explained by the composition of the source water used for irrigation of corresponding grapes.

Table 1. Wine values obtained from NMR spectra

Sample No	Water fraction in wine, vol. %	Ethanol, vol. %	Glycerol, g/l	Relative deuterium amounts in the water fraction of wine, ppm
2	87.17	12.83	6.67	149.1
3	86.91	13.09	5.37	166.4
4	86.64	13.36	8.60	165.6
5	87.42	12.58	5.41	165.6
6	87.17	12.83	5.76	151.7
7	85.76	14.24	8.53	166.3
8	87.06	12.94	7.85	166.2
9	85.30	14.70	8.44	162.4
10	86.58	13.42	8.12	159.1
11	88.32	11.68	5.26	164.2
12	86.06	13.94	8.64	163.8
13	84.75	15.25	10.01	161.9
14	86.70	13.30	8.78	167.6

NMR SPECTROSCOPY OF DEUTERIUM IN ETHANOL OF WINES AND SIMILAR PRODUCTS

The ^2H NMR method was most widely known and applied in conducting evaluations of the deuterium content in the methyl $(\text{D}/\text{H})_{\text{I}}$ and methylene $(\text{D}/\text{H})_{\text{II}}$ groups of the ethanol molecule in order to determine its botanical origin.

The characterisation of WOS by the content of deuterium in fragments of ethanol molecules of wine requires its quantitative isolation. For this purpose, a procedure has been developed in detail, since in different countries it is permitted to add exogenous sugars to the grape must, also from the plants of other botanical species. In 1990, the International Organisation of Vine and Wine (OIV) adopted for the first time the SNIF-NMR method [9] described in [8] as official in the European Union. To date, a significant base of scientific knowledge on the intervals of the content of deuterium in the methyl $(\text{D}/\text{H})_{\text{I}}$ and methylene $(\text{D}/\text{H})_{\text{II}}$ groups of the ethanol molecule of grape origin has been accumulated according to the results of foreign researches. Taking into account agroclimatic (geographical) and technological factors, the minimum and maximum levels for the $(\text{D}/\text{H})_{\text{I}}$ in grape

ethanol are 98.0 and 106 ppm respectively, and for the $(\text{D}/\text{H})_{\text{II}}$ – 121.0 and 136.0 ppm respectively. These ranges also include ethanol extracted from certain other fruits but not sugar beet. Sugar beet belongs like grape to the group of plants of the photosynthesis C3 pathway) and plants of the photosynthesis C4 pathway (e.g. from corn, sugarcane, sorghum and others) [10–11].

The SNIF-NMR method requires the use of NMR spectrometers, with an operating frequency of at least 400 MHz for ^1H nuclei and equipped with a 10 mm tube sensor and a NMR signal stabilisation system ^{19}F .

The Research and Educational Centre of the Peoples' Friendship University of Russia uses a modified SNIF-NMR method based on the author's development [12] for the standard NMR spectrometers to study the botanical origin of ethanol contained in wines and similar products.

One of the features of the developed method is that the stabilisation of the resonance conditions with the ^{19}F signal is not used. Therefore, the measurement of the deuterium content in CH_2 - $(\text{D}/\text{H})_{\text{II}}$ и CH_3 - $(\text{D}/\text{H})_{\text{I}}$ fragments of ethanol molecules will have significant errors in height of the corresponding peaks due to a small drift of the signals. However, this practically does not

Fig.3. The content of deuterium in the water fraction of the red wine samples of the 2015 harvest

Fig.4. NMR spectrometer JEOL JNM-ECA 600 (Japan) with an operating frequency for the ^1H nuclei of 600 MHz and for the ^2H nuclei of 92.102 MHz. Researcher of Scientific Research and Education Centre of the RUDN University – Vassily Ivlev

affect the areas which are measured in the quantitative NMR spectroscopy by the integral but not by the height intensity. A distinct advantage of this approach over the method [9] is that the deuterium content in the hydroxyl group (OH) of extracted ethanol or residual water can be measured by the intensity of the signals which, due to exchange processes, always have a larger width than deuterium (^2H) of the methyl CH_3 (D/H)_I and methylene CH_2 (D/H)_{II} groups. Since the deuterium content in the hydroxyl group is exclusively influenced by the water fraction of the wine, there is no need for a detailed analysis of the ^1H NMR spectrum. It is sufficient to measure the density of the analyte solution to estimate the content of residual water in ethanol containing distillate. Another feature of the developed method is the use of dehydrated dimethylsulfoxide (DMSO) as a calibration additive with a deuterium content increased by two orders of magnitude. This reduces in 10 times (from 30 vol. % in the case of the method [9] to 3 vol. %) reduces the volume fraction of the calibration substance and thereby halves the time of recording the NMR spectrum. The total error of this method ensures the precision of the measured values $\pm 3\%$, which is quite enough to estimate the botanical origin of ethanol.

The method was developed and tested on a JEOL JNM-ECA 600 (Japan) spectrometer with an operating frequency for ^1H nuclei 600 MHz and ^2H nuclei 92.102 MHz. The device is equipped with an autosampler and has 5 mm and 10 mm sensors (Fig.4). Calibrated tubes

with a diameter of (4.97 ± 0.013) mm and a length of 178 mm were used in the research. Sodium trimethylsilyl propionate- d_4 (TSP) was used as an internal standard for the quantitative determination of components to record NMR ^1H spectra, the signal of methyl protons of which does not overlap with the peaks of the analytes. When NMR ^2H spectra are recorded as an internal standard (3 vol. %), DMSO was used with an artificially increased deuterium content by two orders of magnitude. A typical spectrum is shown in Fig.5. The following experiment conditions were selected for recording the spectra on ^2H nuclei: pulse 45° , gain 72, shift 5 ppm, scan 11 ppm, 16K points per spectrum, observation time of free-induction decay 2 s, scan number 800, the delay time between pulses is $5 T_1 (^2\text{D}_2\text{O}) \geq 5$ s. The automatic correction of the baseline, manual phase adjustment and exponential multiplication by the value 2.0 were used to process the spectrum. The table 1 presents the averaged data of five measurements of the deuterium content ^2H in each sample, for which the relative standard deviation is 2%.

CONCLUSION

A new method of a cumulative NMR screening analysis (CS-NMR) was proposed to determine the deuterium content in water of various water-organic solutions including wine and similar products, juices etc. The first stage is the recording of quantitative NMR ^1H spectra to determine the water content in the sample by taking into account other identified components of the solution.

Fig.5. The NMR ^2H spectrum of an alcohol distillate sample with the addition of DMSO: 1 – OH (hydroxyl group of ethanol), 2 – H_2O (residual water in distillate), 3 – CH_2 (methylene group of ethanol ($\text{D}/\text{H}_{\text{II}}$)), 4 – calibration compound DMSO, 5 – CH_3 (methyl group of ethanol ($\text{D}/\text{H}_{\text{I}}$))

Then, the NMR ^2H spectrum is calculated from which the isotope ^1H of water as an important characteristic of origin or authenticity with a control sample. A method that does not require sample preparation is convenient for rapid screening of many water-organic objects. Detailed determination of the isotopic composition of hydrogen of other components of WOS after their isolation can be performed with a standard equipment by a faster original technique.

A modified NMR method is used at the Research and Educational Centre of the Peoples' Friendship University of Russia, to study the botanical origin of ethanol in wine and other alcohol-containing products which is based on the determination of the deuterium content in the methyl ($\text{D}/\text{H}_{\text{I}}$) and methylene ($\text{D}/\text{H}_{\text{II}}$) groups of an alcohol molecule. The quantitative data on the content of deuterium in the methyl group ($\text{D}/\text{H}_{\text{I}}$) of alcohol make it possible to draw a conclusion not only about the botanical form of the plant, but also about the geographic region of its origin, as well as about the nature of the water used in the fermentation of carbohydrates and formation of ethanol. The data on deuterium content in the methylene group ($\text{D}/\text{H}_{\text{II}}$) alcohol indicate the climatic conditions of the growth of the plant (e.g. grapes).

REFERENCES

1. Hachey D.I., Wong W.W., Boutton T.W., Klein P.D. Isotope ratio measurements in nutrition and

biomedical research // Mass Spectrometry Reviews. 1987. V. 6. P. 289–328. DOI: 10.1002/mas.1280060203

2. Scrimgeour C.M., Rollo M.M., Mudambo S.M., Handley L.I., Prosser S.J. A simplified method for deuterium/hydrogen isotope ratio measurements on water samples of biological origin // Biological Mass Spectrometry. 1993. 22. P. 383–387. DOI: 10.1002/bms.1200220704
3. Ma X., Deng P., Wang X., Zhang C., Hou X. Direct determination of deuterium of wide concentration range in water by Nuclear Magnetic Resonance // Talanta. 2012. 97. P. 450–455.
4. Dzhimak S.S., Basov A.A., Fedulova L.V., Bykov I.M., Ivlev V.A., Melkonyan K.I., Timakov A.A. Determination of the deuterium concentration in biological fluids by NMR spectroscopy // Aerospace and Environmental Medicine. 2016. 50 (3). P. 42–47.
5. Varcacel M., Caerdenas S., Gallego M. Sample screening systems in analytical chemistry // Trends in Analytical Chemistry. 1999. 18(11). P. 685–694.
6. Kalabin G.A., Kozlov Yu.P., Rykov R.S., Kulagina N.V., Kushnarev D.F., Rokhin A.V. A new algorithm for identifying natural Georgia wines by NMR spectroscopy // Bulletin of the Peoples' Friendship University of Russia. Ecology and Life Safety Series. 2003. 9. C. 77–82.
7. Christoph N., Hermann A., Wachter H. 25 Years authentication of wine with stable isotope analysis in the European Union – Review and outlook // BIO Web of Conferences. 2015. 5. 02020. 8 p.
8. Martin G.J., Martin M.L., Mabon F., Michon M.J. Identification of the origin of natural alcohols by natural abundance hydrogen-2 nuclear magnetic resonance // Analytical Chemistry. 1982. 54. P. 2380–2382.
9. EC-Regulation 2676/1990 // Official Journal of the European Community. 1990. 64. L272.
10. Bauer-Christoph C., Wachter H., Christoph N., Rossmann A., Adam L. Assignment of raw material and authentication of spirits by gas chromatography, hydrogen- and carbon-isotope ratio measurements. I. Analytical methods and results of a study of commercial products // Zeitschrift fuer Lebensmitteluntersuchung und-forschung A. 1997. 204. P. 445–452.
11. Perini M., Guzzon R., Simoni M., Malacarne M., Larcher R., Camin F. The effect of stopping alcoholic fermentation on the variability of H, C and O stable isotope ratios of ethanol // Food Control. 2014. 40. P. 368–373.
12. Kalabin G.A., Kulagina N.V., Rykov R.S., Kozlov Yu.P., Kushnarev D.F., Rokhin A.V. Identification of the raw material nature of ethanol by the NMR spectroscopy method // Bulletin of the Peoples' Friendship University of Russia. Ecology and Life Safety Series. 2003. 7. P. 87–98.